

Project Specifications and Pre-qualification Information

Project: OR 3 Surgical Lights Replacement

GENERAL REQUIREMENTS:

1. Bidder must have 5 years of OSHPD experience.

TECHNICAL/INSTALL SPECIFICATIONS:

1. Include all infection control containments for the project. Reference the attached Infection Control Permit for requirements.
2. Bidder to provide all conduits for power, data, and video as shown on Contract Drawings.
3. Bidder to provide and pull power and data cables. Video cables will be provided and pulled by Stryker.
4. Bidder to make all electrical connections.
5. Contractor to provide steel supports and patch drywall (all work above ceiling). Lights will be installed by Stryker.
6. Owner will provide a waste dumpster onsite.
7. Bidder to repaint all the walls and ceilings (match existing paint colors).